

GOVERNMENT OF N.C.T. OF DELHI
ARUNA ASAF ALI GOVERNMENT HOSPITAL
5, RAJPUR ROAD, DELHI-110054
REPAIR & MAINTENANCE BRANCH

File No-F.8(11)/R&M/2021/AAAGH/ 1283

Dated:- 24/10/2021

TO,

- 1) M/S EDGE MEDICAL SOLUTIONS PRIVATE LIMITED, add -- HOUSE NO 104 ,FIRST FLOOR ,ROHINI SEC- 2 VILLAGE MANGOLPUR KALAN WATER TANK , New Delhi --110085
- 2) DRS CONSULTANTS PRIVATE LIMITED, Add- WZ -237, SHAKURPUR VILLAGE POST OFFICE Delhi - 110034
- 3) TRIVENI BIOTRONIX PRIVATE LIMITED , Add- PLOT NO 192, GALI NO 8, DEEP ENCLAVE 1, VIKAS NAGAR, UTTAM NAGAR , New Delhi , 110059
- 4) BIONATION IMAGING PRIVATE LIMITED, Add- PLOT NO 192A , GALI NO -8 ,DEEP ENCLAVE 1 ,VIKAS NAGAR UTTAM NAGAR , New Delhi – 110059.
- 5) RAYDENT MEDITECH PRIVATE LIMITED, Add- H NO 552 KOTLA VIHAR, PHASE 2 NANGLOI ,NEAR FARM HOUSE , Delhi- 110043

SUBJECT :- Inviting the sealed quotations for Quality Assurance (QA) TEST report of portable X-ray machine as per AERB guidelines - regarding.

Sir/Madam

This Hospital is in process to get the sealed quotations for Quality Assurance (QA) TEST report of portable X-ray machine as per AERB guidelines- regarding from AERB Authorized / Approved firms.

Details of machine :-Make/Model- M/S. Skanray technologies pvt. ltd , mysore , model- Skanmobile , Sr. no. - MY5218R0025 available in the Radiology dept of AAAGH which was installed on 28/03/2019 in this hospital and is functional and also under CAMC till 19/12/2025.

Sealed Quotations by all Eligible Firm in this regard are invited **within Two weeks of dispatch of this letter / closing date on website of Hospital/ H&FW Dept, GNCTD whichever occurs later,** in the Tender Box lying on 1st floor of Administrative Block of this hospital.

This is issued with the prior Approval of worthy Medical Superintendent..

/

In-charge – Repair & Maintenance Branch
AAAGH, New Delhi

File No-F.8(11)/R&M/2021/AAAGH/ 1283
Copy for information and necessary action to

Dated:

- ✓ 1- PS to MS
2- I/C- Dept of Radiology


In charge – Repair & Maintenance Branch
AAAGH, New Delhi